




Part **Beaker**
 Part Number **03061-2**
 Serial Number **93801 - 93900**



Test Certificate

| | | | USP / Ph.Eur. Tolerances | |
|-------------------|--------------------|----------------------|-------------------------------------|--------------|
| Dimensions | Beaker | Inside Ø a | 104.5 to 107.5 mm | 97 – 115 mm |
| | | Height b | 145 mm | 138 – 160 mm |
| | | With straight bottom | | |
| Material | Borosilicate glass | | | |

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below) and the requirements of ISO 3819:2015..

Aesch, 2023-08-08

Michael Schmidlin
 SOTAX Quality Management
 SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland